



HGSC Clinical Laboratory
 One Baylor Plaza, Houston TX, 77030
 Phone: 713.798.6539 Fax: 713.798.5741
www.hgsc.bcm.edu questions@hgsc.bcm.edu



Patient Name:	Joe Nguyen	Sample Collected Date:	11/20/2015
Patient ID:	NA19202	Sample Received Date:	12/02/2015
Age:	41	Report Date:	1/12/2016
DOB:	3/7/1975	Sample Type:	DNA
Sex:	Male	Indication for Testing:	None
Patient Sample ID:	A4F0G7H5J2	Physician Info:	Dr. Yun Choi
Accession #:	0987654321C		



eMERGE III CAREseq Version 2 Sequencing Report

This test interrogates the protein-coding and exon-splicing regions of 109 genes as well as 1551 single-nucleotide polymorphisms that may impact human health and disease.

Pathogenic Variants

1. None Detected

Table 1: Pathogenic Variant Details

None Detected

Likely Pathogenic Variants

1. None Detected

Table 2: Likely Pathogenic Variants Details

None Detected

Comments & Recommendations:

No variants that meet reporting criteria were identified in this sample.

Gene Coverage:

All genes have 100% of targeted bases sequenced to redundant coverage of 20x or greater with the following exceptions: CACNA1B (96.94%), APOB (99.39%), COL5A1 (98.03%), RYR1 (98.74%), KCNQ1 (97.7%), TGFB1 (93.56%).

Further information is available in the ExCID report.

Methodology:

1. CAREseq Version 2 NGS Panel*: for the paired-end pre-capture library procedure, genome DNA is fragmented by sonicating genome DNA and ligating to the Illumina multiplexing PE adapters (reference 1). The adapter-ligated DNA is then PCR amplified using primers with sequencing barcodes (indexes). For target enrichment capture procedure, the pre-capture library is enriched by hybridizing to biotin labeled in-solution probes (reference 2) at 56°C for 16 - 19 hours. For massively parallel sequencing, the post-capture library DNA is subjected to sequence analysis on Illumina HiSeq platform for 100 bp paired-end reads. The following quality control metrics of the sequencing data are generally achieved: >70% of reads aligned to target, >99% target base covered at >20X, >98% target base covered at >40X, average coverage of target bases >200X. SNP concordance to SNPTrace genotype array: >99%. This test may not provide detection of certain genes or portions of certain genes due to local sequence characteristics or the presence of closely related pseudogenes. Gross deletions or duplications, changes from repetitive sequences may not be accurately identified by this methodology.

2. As a quality control measure**, the individual's DNA is also analyzed by a SNP-array (Fluidigm SNPTrace panel (reference 3)). The SNP data are compared with the NGS panel data to ensure correct sample identification and to assess sequencing quality.

3. Data analysis and interpretation by Mercury 3.4 (reference 3): The output data from Illumina HiSeq are converted from bcl file to FastQ file by Illumina bcl2fastq 1.8.3 software, and mapped by the BWA program (reference 4). The variant calls are performed using Atlas-SNP and Atlas-indel developed in-house by BCM HGSC. Variant annotations are performed using the Cassandra tool, developed in-house. Synonymous variants, intronic variants not affecting splicing site, and common benign variants are excluded from interpretation unless they were previously reported as pathogenic variants. The variants were interpreted according to ACMG guidelines (reference 5) and patient phenotypes. Variants related to patient phenotypes are usually confirmed by Sanger sequencing for patients and if available, parents. Sanger confirmation is noted in the "Notes" section of the tables if performed. It should be noted that the interpretation of the data is based on our current understanding of genes and variants at the time of reporting.

References:

1. Illumina, Inc. (2011) Multiplexing Sample Preparation Guide (Part # 1005361 Rev. D). 2011.
2. Roche NimbleGen, Inc. (2010) NimbleGen SeqCap EZ Exome Library SR User's Guide (Version 2.2).
3. Liang-Chu MM, Yu M, Haverty PM, Koeman J, Ziegler J, Lee M, Bourgon R, Neve RM. Human biosample authentication using the high-throughput, cost-effective SNPTrace(TM) system. PLoS One. 2015 Feb 25;10(2):e0116218.
4. Li H. and Durbin R. (2009) Fast and accurate short read alignment with Burrows-Wheeler Transform. Bioinformatics, 25:1754-60. PMID:19451168.
5. Richards S, Aziz N, Bale S, Bick D, Das S, Gastier-Foster J, Grody WW, Hegde M, Lyon E, Spector E, Voelkerding K, Rehm HL; ACMG Laboratory Quality Assurance Committee. Standards and guidelines for the interpretation of sequence variants: a joint consensus recommendation of the American College of Medical Genetics and Genomics and the Association for Molecular Pathology. Genetics in Medicine (2015) 17, 405-423. PMID: 25741868.

HGSC Clinical Laboratory – Molecular Diagnostics Report		
Patient Name: Joe Nguyen	Patient ID: NA19202	Accession #: 0987654321C

Christine M. Eng, M.D.
Medical Director
May 4, 2016

Yaping Yang, Ph.D
Laboratory Co-Director
May 4, 2016

This test was developed and its performance determined by this laboratory. It has not been cleared or approved by U.S. Food and Drug Administration. Since FDA is not required for clinical use of this test, this laboratory has established and validated the test's accuracy and precision, pursuant to the requirement of CLIA '88. This laboratory is licensed and/or accredited under CLIA and CAP (CAP# 8004250 / CLIA# 45D2027450).